Laser and IPL Technology in Dermatology and Aesthetic Medicine



Laser and Christian Raulin Syrus Karsai Editors IPL Technology in Dermatology and Aesthetic Medicine



Editors Christian Raulin and Syrus Karsai

Laser and IPL Technology in Dermatology and Aesthetic Medicine



Editors

Christian Raulin Laserklinik Karlsruhe, Kaiserstr. 104, Karlsruhe, 76133 Karlsruhe, Germany info@raulin.de

Syrus Karsai Laserklinik Karlsruhe, Kaiserstr. 104, Karlsruhe, 76133, Germany info@raulin.de

ISBN 978-3-642-03437-4 e-ISBN 978-3-642-03438-1

Springer Heidelberg Dordrecht London New York

Library of Congress Control Number: 2010937978

© Springer-Verlag Berlin Heidelberg 2011

This work is subject to copyright. All rights are reserved, whether the whole or part of the material is concerned, specifically the rights of translation, reprinting, reuse of illustrations, recitation, broadcasting, reproduction on microfilm or in any other way, and storage in data banks. Duplication of this publication or parts thereof is permitted only under the provisions of the German Copyright Law of September 9, 1965, in its current version, and permission for use must always be obtained from Springer. Violations are liable to prosecution under the German Copyright Law.

The use of general descriptive names, registered names, trademarks, etc. in this publication does not imply, even in the absence of a specific statement, that such names are exempt from the relevant protective laws and regulations and therefore free for general use.

Product liability: The publishers cannot guarantee the accuracy of any information about dosage and application contained in this book. In every

individual case the user must check such information by consulting the relevant literature.

Cover design: Cover design: eStudioCalamar, Figueres/Berlin

Printed on acid-free paper

Springer is part of Springer Science+Business Media (www.springer.com)

Foreword I

At the publication of this volume, lasers are half a century old, or perhaps it is more accurate to write half a century young. The interplay between light and matter fascinated Albert Einstein, whose $E = mc^2$ expresses the fact that all matter and all energy are equated by only one constant – the speed of light. He also figured out that the only "thing" different observers measure as being constant – is the speed of light. Apparently, we live in a universe not only bathed in light, but defined by it. Einstein's work on quantum electromagnetism is less well-known, but anyone using a laser should appreciate that the device exists because of it. Fundamentally, creation of a photon is the time-inverse of its absorption. In an excited atom, it is possible for photons to stimulate the creation of more photons. Stimulated emission, the S.E. of L.A.S.E.R., is at the heart of every laser. Charles Townes and colleagues made the first device using stimulated emission to amplify (the "A" in L.A.S.E.R.) microwave photons, and in a few years this was done with light (the "L" in L.A.S.E.R.).

Is an understanding of laser operation important to a physician using a laser? Yes! Extreme power, coherence, well-defined wavelength, ability to generate very brief pulses, and to focus tightly, are all properties of lasers important to their medical applications. In dermatology, we use lasers mainly as surgical tools to heat or vaporize cells and tissue with precision and often with selectivity, i.e., preferentially affecting some light-absorbing targets within the skin. This excellent book is a thorough and practical description of those applications, which depend on high-power pulses at well-defined wavelength regions. Sometimes, a xenon flashlamp can be used for selective photothermolysis of large targets such as hair follicles or blood vessels. The selectivity possible with flashlamps is always inferior to that possible with lasers, but in practice, is often sufficient. For very small targets such as the isolated dermal melanocytes of nevus of Ota, or tattoo ink-containing cells, very powerful short pulses are needed to achieve safety and efficacy. Lasers are the only sources able to deliver such pulses.

In other applications, such as fractional photothermolysis, it is necessary to deliver light in a microbeam, or as an array of small foci. Again, lasers are ideal for this application. Fractional laser treatments at present are not selective for some specific target in the skin. However, we can expect that "smart" laser microbeam treatments will be developed that will use imaging to identify which tissue structures the laser should expose. These devices may create another paradigm for laser dermatology, because some of the most important target structures in skin are not pigmented.

The coherence property of lasers is not used much in dermatology, but it will be. Coherence can be thought of as a measure of how much alike photons in a beam are, with respect to each other. High coherence lasers are used to make holograms, and in medicine to measure blood flow by laser doppler velocimetry. Lasers have revolutionized biological microscopy for research. They will probably revolutionize microscopy of living skin, with further development. A combination of high brightness and low coherence offers the potential for deep, high-speed microscopy of skin. Optical coherence tomography and related approaches used widely in ophthalmology, are entering cardiovascular and endoscopic imaging applications, and have strong potential in dermatology.

Ultimately, it is clinical practice that defines the therapeutic and diagnostic use of lasers. As such, clinical dermatologists have a major role to play in motivating the creation of new applications and new medical lasers. This book offers practical knowledge, practical uses, and a thorough understanding of the risks and benefits posed by a wide spectrum of medical lasers and related technologies.

People are not pastries. If you treat them according to a fixed recipe, they will sometimes be overcooked. The natural variations of skin structure, pigmentation, sun exposure, wound healing, age, and gender are important for not only laser dosimetry, but also for the fundamental decisions of whether to treat and if so, with which device(s). These are also the natural variations that make medicine an art, as well as a science. Enjoy!

R. Rox Anderson Boston, MA, USA

Foreword II

When amplified by stimulated emission of radiation, light may have unique properties capable of causing novel reactions in biological materials. In complex multicellular organisms, the host responses to this form of electromagnetic radiation may also be unique resulting in outcomes different from those expected from traditionally described response to injury.

Until the twentieth century, the complex forces guiding the evolution of human responses to injury were never influenced by the collective properties of lasers or IPL technology. By altering the properties of lasers, savvy investigators have explored the creation of selective tissue injury with novel combinations of spatial localization and mechanisms of tissue alteration. The form of injury may be manipulated to stimulate a cascade of repair responses aimed at correcting a variety of lesions in skin and other organs. The precision of tissue alteration can not be achieved by other means. This book contains many examples – written by those who discovered the response pathway and those who have successfully exploited the tissue response to develop and constantly improve the diagnostic and therapeutic use of lasers and IPL technologies.

We expect more rapid growth of the technologies that creates unique energy waves, pulses, and "packets" of energy. We expect advances in molecular biology, pathology, optical diagnostics, and molecular tagging to lead a better understanding of tissue response to radiation. We expect investigators versed in biology, photochemistry, nanotechnology, electronics, and photophysics to continue the rapid development of this technology in subsequent medical and surgical uses.

There will always be those creative investigators who understand both the technical and biological parameters enough to capture selected portions of tissue responses at the right place at the right time in the right organ.

This book may be viewed as an update in the impression progress made to date and a foundation for future advances.

> John A. Parrish Boston, MA, USA

Preface

Laser and IPL technology have brought about thorough ongoing changes in the treatment options for many clinical pictures encountered in dermatology during recent years. The specialism of aesthetic medicine has benefited greatly from all this. The rapid technical progress and the frequent emergence of new equipment and indications make it extraordinarily difficult for the interested doctor to get an overview of the research position and the possibilities and the limitations of these procedures. Disconcertingly, popular science publications also have a role in all this, so that the need for factual and objective information is constantly becoming more intense, especially as there are very few evidence-based scientific studies available on the possible applications of new laser techniques. In addition, the time elapsing between current knowledge and developments and their practical implementation is becoming shorter and shorter.

This book is addressed equally to specialists who already have experience in using laser and IPL systems and to doctors and students who harbor the ambition of becoming well-informed about the specific treatment options so as to be in a position to give patients competent advice. Without a precise and structured digest of results from research and technology, including the most recent, it is not possible for the clinician to guarantee upto-date and correct care of the patient.

We have been able to secure international opinion of leaders for all the chapters of this book. This was the only way to fulfill our own ambition of taking full account of the rapid further development of the specialist area and to give the reader an up-to-date overview of laser therapy in dermatological-aesthetic medicine and its fringe areas, including photodynamic therapy. We have tried consistently to achieve a uniform presentation, with reference also to terminology and the structure of the texts. Where there turned out to be differences in various authors' opinions, both sides had the chance to make their case.

This textbook gives the theoretical bases of laser and IPL technology and reviews the voluminous current knowledge on the various indications. References to unsolved problems or questions that are still under discussion are found in the appropriate text sections. The chapters each have a list of the background literature and further reading matter. It would not have been possible to avoid overlappings and repeats, and they are intentional to ensure that each of the individual texts can be complete in itself.

We are particularly concerned to profile conventional alternatives to laser treatment insofar as they are better based in medicine and more economical. The quality of a method is determined to a decisive degree by careful observation of indications, and laser and IPL technology are far from being a universal panacea – not least because of the potential risks inherent in them. We therefore thought it was important to sensitize the reader to complications, safety, administration of treatment by persons with no medical qualifications, and forensic considerations.

This book is no more able than any other to be a substitute for first-hand experience with patients. Consultation and critical analysis of diagnosis and therapy options remain an essential component of the doctor's work. For this reason one chapter is devoted entirely to ethical aspects of laser therapy.

Our sincere thanks are due to all the authors for their committed cooperation on this multiauthor work. We thank our staff assistants and colleagues for their constructive criticisms and their many hours of hard work without which it would not have been possible to prepare the book within the planned time frame. All patients who are recognizable in the illustrations have kindly given their consent to this. At this point, we should like to thank them cordially for their support. Our book is not made up solely of theory, but takes life from the many illustrative clinical examples. Not least, we are very grateful to Springer-Verlag, especially the clinical medicine editors Dr Tobias Kemme, Dr Sverre Klemp and Mrs Ellen Blasig, for their efficient project management and the successful structure of the book, and Mr Sreejith Viswanathan for his excellent editorial work.

We are pleased to say we feel confident that this text has presented laser and IPL therapy as used in dermatology and aesthetic medicine in a wellrounded manner and on the basis of comprehensive and up-to-date knowledge of the current scientific status.

> Christian Raulin Syrus Karsai Karlsruhe autumn 2010

Contents

Part I Basics

- 1 Basic Laser Physics Rudolf Steiner
- 2 Laser-Tissue Interactions Rudolf Steiner
- **3 Intense Pulsed Light Technology** Karin Kunzi-Rapp and Rudolf Steiner

Part II Indications

- 4 Benign Tumors and Organoid Nevi Stefan Hammes
- **5** Facial Rejuvenation and Other Clinical Applications of Intense Pulsed Light

Peter Bjerring and Kaare Christiansen

6 Inflammatory Dermatoses: Acne Mathew W. Ludgate and Jeffrey S. Orringer

7 Inflammatory Dermatoses: Other than Acne Vulgaris

Pablo Boixeda, Lorea Bagazgoitia, João Borges da Costa and Maria Calvo

8 Laser-Assisted Liposuction

Boris Sommer and Dorothee Bergfeld

9 Photoepilation of Unwanted Hair Growth

Annesofie Faurschou and Merete Haedersdal

10 Pigmentation and Hypopigmentation: Benign Pigmented Lesions

Henry H. L. Chan

- **11 Pigmentation and Hypopigmentation: Vitiligo and Other Disorders** Thierry Passeron and Jean-Paul Ortonne
- **12 Precancerous Conditions and Malignant Tumors** Christian Raulin, Syrus Karsai and Laurenz Schmitt
- **13 Scars, Keloids, and Stretch Marks** Hilda Justiniano, Andrea Willey and Suzanne L. Kilmer
- **14 Laser Treatment of Tattoos and Other Dyschromia** Syrus Karsai and Christian Raulin
- **15 Varicose Veins: Endovenous Laser Treatment** Serge R. Mordon and Marc E. Vuylsteke
- **16 Common Vascular Lesions** Emil A. Tanghetti, Mirko Mirkov and Rafael A. Sierra
- **17 Vascular Lesions: Port Wine Stains** Sean Lanigan
- **18 Vascular Lesions: Hemangiomas** Margitta Poetke, Peter Urban and Hans-Peter Berlien
- **19 Rare Vascular Lesions** Sean Lanigan

20 Viral Infections Daisy Kopera

21 Wrinkles and Acne Scars: Ablative and Nonablative Facial Resurfacing

Melissa A. Bogle, Geeta Yadav, Kenneth A. Arndt and Jeffrey S. Dover

22 Wrinkles and Acne Scars: Technology-Based Treatment of Periorbital Wrinkles Laurel Naversen Geraghty and Brian S. Biesman

- **23 Wrinkles and Acne Scars: Fractional Ablative Lasers** Arielle N. B. Kauvar and Melanie A. Warycha
- 24 Wrinkles and Acne Scars: Fractional Nonablative Lasers Uwe Paasch

Part III Special Aspects

- **25 Complications in Laser Surgery and IPL Treatment** Christian Raulin, Syrus Karsai and Laurenz Schmitt
- **26 Cooling Techniques** Stefan Hammes
- 27 Ethical Considerations in Aesthetic Medicine Giovanni Maio
- **28 Photodynamic Therapy** Philipp Babilas and Rolf-Markus Szeimies
- **29 Quality Standards in Aesthetic Medicine** Hans-Robert Metelmann, Peter D. Waite and Stefan Hammes
- **30 Safety/Eye Protection** Wolfgang Bäumler
- Part IV Afterword
- **31 The Business of Laser Surgery** Christopher B. Zachary
- **32 The Hall of Shame** Christopher B. Zachary

Index

Part 1 Basics

1. Basic Laser Physics

Rudolf Steiner¹[™]

(1) Institut für Lasertechnologien in der Medizin und Meßtechnik an der Universität Ulm, Helmholtzstraße 12, 89081 Ulm, Germany

🖾 Rudolf Steiner

Email: rudolf.steiner@ilm.uni-ulm.de

Abstract

Following the history of lasers and, in parallel, the applications in medicine demonstrates the continuous improvement of insight into laser reactions and development of devices. Understanding the characteristics of photons and light is the basis for all applications in medical diagnosis and therapy. Knowledge about the physical principle of lasers will convince you that there is no mystery behind lasers. Lasers can be designed in different ways and thus be optimised to the intended use. Laser radiation, depending on the wavelength and the mode of operation, is a powerful tool to treat patients, but it also can be dangerous if safety instructions are ignored. An overview about the different types of medical lasers actually on the market facilitates the right choice of a laser system for the specific application.

1.1 Core Messages

- Following the history of lasers and, at the same time, their medical applications demonstrates the continuous improvement of insight into laser reactions and the development of devices.
- Understanding the characteristics of photons and light is the basis for all applications in medical diagnosis and therapy.
- Knowledge about the physical principle of lasers will remove any mystery from laser technology.

- Lasers can be designed in different ways and thus be optimised to the intended use.
- Laser radiation, depending on the wavelength and the mode of operation, is a powerful tool to treat patients, but it also can be dangerous if safety instructions are ignored.
- An overview about the different types of medical lasers on the market facilitates the right choice of a laser system for the specific application.

1.2 History

To invent the LASER (acronym for *L*ight *A*mplification by *S*timulated *E*mission of *R*adiation, coined by Gordon Gould) some understanding of Einstein's ideas about stimulated emission of radiation [5] was necessary. Albert Einstein developed the theoretical concept of light travelling in waves of particles (photons) and of "stimulated emission." Although Einstein did not invent the laser his work laid the foundation for it. He pointed out, considering photon statistics, that stimulated emission of radiation could occur (Fig. 1.1).

The precursor to the laser was the MASER (microwave amplification by stimulated emission of radiation). The impetus for the development of the maser was increasing interest in microwave radiation following the utility of it in radar technology. The first maser [22] was created by Townes (published in 1954) who, along with James Gordon and Herbert Zeiger, succeeded in producing an inverted population by isolating excited ammonia molecules. But the first maser, a system with two energy levels, was incapable of continuous output. To achieve continuous output, new systems with more than two energy levels had to be designed. These systems could release stimulated emission of excited atoms without falling to the ground state, thus maintaining a population inversion. Nikolai Basov and Alexander Prokhorov in Moscow first developed this idea. Basov, Prokhorov, and Townes shared the 1964 Nobel Prize in physics for developing the maser concept (Fig. 1.2).

After masers became a reality, scientists looked at the possibility of stimulated emission in other regions of the electromagnetic spectrum. Arthur Schawlow, together with Townes, at Bell Telephone Laboratories in Murray Hill, New Jersey, began investigating the possibility of optical and infrared (IR) masers. They published in 1958 the first detailed proposal for building an optical maser (later to be renamed a laser) in *Physical Review* [10]. The challenges of creating a working laser were tremendous. The much smaller wavelengths of visible light and the difficulty of finding an appropriate laser medium meant that many more experiments had to be performed and that it was much more difficult

to adjust an apparatus to function as a laser. It was not until 1960 that Maiman [14, 24] created the first working laser. Maiman's laser was a ruby rod with silver ends that had been placed inside a spring-shaped flashlamp. Maiman's laser, however, was only capable of pulsed operation due to its three energy level transitions. Soon afterwards, in 1960, Peter Sorokin and Mirek Stevenson at IBM Laboratories developed the first four-level laser, which was capable, in theory, of continuous output; however, in solid state a continuous output could not be achieved. Further laser developments followed in a short time: the Nd:YAG laser, CO_2 laser, Argon-ion laser, Excimer laser, and diode laser. A list of milestones in the development of lasers is given in Table 1.1.

Table 1.1 Milestones in the history of lasers

1917	Einstein publishes his paper about stimulated emission		
1951	Development of the maser by C.H. Townes		
1958	C.H. Townes and A.L. Schawlow propose that the maser concept could be extended to optical frequencies		
1960	T.H. Maiman at Hughes Labs and his assistant Charles Asawa report on May 16 the first functioning laser: a pulsed ruby laser		
1961	First medical laser applications in ophthalmology by Charles J. Campbell		
1961	The first continuous wave laser is reported by A. Javan: the helium neon laser		
1962	First diode laser is reported by M.I. Nathan et al.		
1963	First medical laser applications in dermatology by L. Goldman		
1964	Nicolay Basov, Charles Townes, and A.M. Prokhorov get the Nobel prize for their fundamental work in the field of quantum electronics, which has led to the construction of oscillators and amplifiers based on the maser and laser principles		
1964	The Argon Laser is developed at Hughes Labs. The continuous wave 488/514 nm argon ion laser is well suited to retinal surgery		
	Kumar Patel invents the CO ₂ (carbon dioxide) 10,064-nm (far-infrared) gas ion laser at Bell Labs		
	The Nd:YAG (neodymium doped, yttrium aluminium garnett) 1,064-nm laser also is developed at Bell Labs		
1969	The dye laser is introduced at IBM Labs by P. Sorokin and J. Lankard. The pulsed dye laser is the first laser to produce selective light-induced injury		
1970	The first excimer laser [based on Xenon (Xe) only] is invented by N. Basov's group at Lebedev Labs, Moscow		
1977	J. Madey's group at Stanford University develops the first free electron laser		
1980	L. Goldman establishes <i>The American Society for Laser Medicine and Surgery, Inc</i> , the world's largest professional organization dedicated to promoting excellence in patient care by advancing laser applications and related technologies		
1981	A. Schawlow and N. Bloembergen receive the Nobel prize for "their contribution to the development of laser spectroscopy and nonlinear optics"		
1984	D. Matthew's group at Lawrence Livermore Labs demonstrates a "laboratory" x-ray laser		

2005 J. Hall and T. Hänsch receive the Nobel prize for contributing to the development of laser-based precision spectroscopy, including the optical frequency comb technique



Fig. 1.1 Albert Einstein (left) and T.H. Maiman (right)





Arthur L. Schawlow (1981) and Theodor W. Hänsch (2005)

Fig. 1.2 Nobel laureates connected with lasers [Alexander M. Prokhorow, Charles Hard Townes, Nicolay G. Basov (1964), Arthur L. Schawlow (1981), and Theodor W. Hänsch (2005)]

1.3 Characteristics of Light

Light is electromagnetic radiation in the visible wavelength region from 400 to 700 nm or extended from 380 to 750 nm. Laser radiation, however, is between 100 nm and 1 mm. The speed of light in a vacuum is $c \text{ (ms}^{-1})$, where

 $c = 299,792,485 \text{ms}^{-1}(or 186,282 \text{miles per second}).$

Speed of light, the frequency v of the electromagnetic wave, and the wavelength λ in a vacuum obey the relation

 $c = \lambda.\nu$ andinmatterc $(\lambda) = c_0/n(\lambda)$

(n: the wavelength dependent refractive index).

The electromagnetic radiation is characterised by Maxwell's equations of an alternating electric field, with field strength E (V/cm) and magnetic field H (Fig. 1.3). The transported power is described by the Poynting vector S (W/m²) representing the energy flux density



Fig. 1.3 Schematic presentation of an electromagnetic wave with electric, E, and magnetic, H, field vectors and the Poynting vector, S. The planes of the linearly polarised field vectors are perpendicular to each other

 $S = E \times H.$

According to Bohr's atomic model, electrons occupy certain fixed orbitals (principal quantum number). Electrons can only jump from orbit to orbit and therefore they can emit energy (emission of radiation) or absorb energy (absorption of radiation). The quantisation of energy is described by ΔE (eV). (Fig. 1.4).

$$\begin{split} \Delta E &= h.\nu(h:Planck' \text{sconstant}; \\ h &= 6.26617 \times 10^{-34} \text{Js}; \nu = frequency of light). \end{split}$$

With this expression, light can be described either as a wave or as a particle (photon) moving at the speed of light, having no mass but carrying energy and momentum. This property is referred to as wave-particle duality [3, 20]



Fig. 1.4 Schematic presentation of the origin of spectral lines of a hydrogen atom

$$\begin{split} Energy: E_{photon} &= h.\nu, \\ Momentum: p &= E_{photon}/c = h.\nu/c, \\ Power: P &= N_{photon}.h\nu/t (N:number of photons). \end{split}$$

1.3.1 Stimulated Emission of Radiation

Photons can be absorbed by atoms when the enegy corresponds to an electron transition to a higher energy level (orbit). The atom is then in an excited state with energy E_2 . It may spontaneously decay to the ground state, with energy E_1 , emitting a photon with the difference in energy between the two states, $\Delta E = E_2 - E_1 = hv$, where h is again Planck's constant. The phase of the spontaneously emitted photon and also the direction of travel are random. This is different for stimulated emission (Figs. 1.5 and 1.6).



Fig. 1.5 Absorption and spontaneous emission. A photon is absorbed by an atom, lifting an electron into a higer orbit. This excited state lasts only for about 10^{-7} s, then the electron drops into a lower orbit with emission of a photon of energy $\Delta E = E_2 - E_1$



stimulated emission

Fig. 1.6 Stimulated emission. A photon induces the decay of an excited energy state. Simultaneously, a second photon is emitted that has the same phase and wavelength

The interaction of a photon with an atom will affect the atom's state. The atom, acting as a small electric dipole, will oscillate with the external field. Thus, an electron, due to the presence of another photon, is influenced to decay to the lower energy level, releasing a second photon that is in phase with the first photon. The second photon also travels in the same direction and has the same energy and polarisation. This is called *stimulated emission*. As a result, the two photons are totally coherent. This is also the basis for optical amplification to take place.

If the number of atoms in the excited state is given by N_2 , the rate at which stimulated emission occurs is given by:

$$dN_2/dt = -B_{21}r(v)N_2,$$

where B_{21} is a proportionality constant for this particular transition and is also called the Einstein B_{21} coefficient. $\rho(v)$ is the radiation density of photons of the frequency v. The emission rate of photons is therefore proportional to the number of excited atoms, N_2 , and the density of the perturbing photons. For spontaneous emission we have $dN_2/dt = -A_{21}N_2$ and for induced absorption we have $dN_1/dt = -B_{12}\rho(v)N_1$.

1.3.2 Definitions of Radiation 1.3.2.1 Radiant Power The power output of a source is described by its *radiant power*, P(W). It is measured with a power or energy metre by photodiode or thermal sensors.

1.3.2.2 Radiant Energy

This is energy emitted, transferred, or received as radiation. A source with a power output P(W), which is turned on for a time duration t(s), will yield a *radiant energy*, Q(J).

Q = P.t(J; Ws).

1.3.2.3 Radiant Intensity

The power P(W) from a point source that is directed into a particular direction at a solid angle, Ω (steradians or sr), is called the *radiant intensity*, I(W/sr).

For an isotropic source, the whole sphere has 4π (sr) of solid angle. In limit of small angle θ , the solid angle of a cone Ω is (Fig. 1.7):



Fig. 1.7 Schematic representation of radiant intensity from a point source

 $\Omega = 4p_{\cdot} pr^2/4pR^2(sr) = pr^2/R^2(sr) = pq^2(sr).$

1.3.2.4 Irradiance

The power P(W) of a continuously emitting source that irradiates a surface area, $A(cm^2)$, is called the *irradiance*, E, where

$$E = P/A(W/cm^2).$$

1.3.2.5 Radiant Exposure

The energy Q(J) of a pulsed light source that reaches a surface $A(cm^2)$ due to an irradiance E, and is maintained for a time interval t(s) is called *radiant* exposure, H, where

 $H = Q/A = Pt/A = Et(J/cm^2).$

1.3.2.6 Fluence Rate

Imagine an isotropic sphere located in biological tissue collecting photons impinging from all angles onto a small region of its surface. Then it measures the *fluence rate*, F (W/cm²). The fluence rate is the total absorbed power by the small sphere divided by its cross-sectional area, A ($A = \pi r^2$). The fluence rate irradiating the target sphere is:

 $F = P_{absorbed} / A(W/cm^2).$

In some literature the fluence rate is expressed with the greek symbol Φ .

1.4 Principle of Lasers

A laser is actually an *oscillator* rather than a simple amplifier. The difference is that an oscillator has positive feedback in addition to the amplifier. "Light" is understood in a general sense of electromagnetic radiation with wavelengths from 100 nm to 1 mm. Thus one can have ultraviolet, visible, or IR lasers. In medical applications nearly all types of lasers are used for specific medical treatments or diagnostic applications.

The amplifier of a laser is the laser material that can be a solid, a gas, or a liquid. The feedback mechanism is produced by the resonator, where the light is reflected by two mirrors so that the photons pass several times through the laser material. The number of photons within the resonator increases exponentially due to the stimulated emission (Fig. 1.8).



Fig. 1.8 Principle of a laser

However, a laser can only work when the stimulated emission of excited atoms of a laser material is larger than the spontaneous emission and the losses within the resonator. In thermal equilibrium, according to the Boltzmann statistics, the number of excited atoms, N_2 , with higher energy, E_2 , is always smaller than the number, N_1 , of atoms in the lower energy level, E_1 . Therefore, a two-level laser system cannot exist because there must be the possibility of an *inversion*. This is possible with a three- or four-level laser design.

The rate balance of a two-level system at thermal equilibrium using the Einstein's coefficients is

```
(Induced absorption) B_{12}\rho(\nu) N_1 = B_{21}\rho(\nu) N_2 \ (stimulated temission) + A_{21}N_2(spontaneous emission).
```

With the relation of the Einstein coefficients $A_{21}/B_{21} = 8\pi hv^3/c^3$, we receive also Planck's law of black body radiation:

$$\begin{split} r(u) = &8phu^3/c^3 \times 1/(\epsilon^{h\nu/kT}-1) \\ Planck/slawofblackbodyradiation. \end{split}$$

The premise for a laser activity is the population inversion between the levels where the laser transition should take place. The higher this inversion, the more effective is the laser. To get a high inversion, the Einstein coefficient A_{21} of the upper laser level should be small so that the losses due to spontaneous emission are small as well. This means that the lifetime of the upper laser level τ_2 should be long. The lower laser level should be less populated (Fig. 1.9) and should have a short lifetime.



Fig. 1.9 In thermal equilibrium (Boltzmann statistics) the population distribution follows an exponential curve, meaning that the number of excited atoms with higher energy is always less than at lower energy levels ($N_2 < N_1$). Population inversion is the opposite ($N_2 > N_1$)

1.4.1 Laser Design

To produce a population inversion in the laser material of a solid-state laser, one needs a pumping source, which can be a flashlamp or another pumping laser (diode laser). The absorption of the ground state must be sufficiently high. A large Einstein coefficient *B* is favourable. These conditions are well fulfilled in three-level and four-level systems (see Fig. 1.10).



Fig. 1.10 Three- and four-level laser systems

The first laser, which was demonstrated and produced by Maiman, was a three-level ruby laser. Ruby is an Al₂O₃ crystal in which about 0.05% of the Al ions are replaced by Cr³⁺ ions. The problem of a three-level laser is that the lower laser level is the ground state, as shown in Fig. 1.10. It is much more difficult to obtain population inversion in three-level lasers because the lower laser level initially has a very large population, N_0 . By turning on the flashlamp pump, dN atoms are excited and lifted to level 1, which then decay quickly to level 2. Thus, the population of level 2 will be dN, and the population of the ground state will be $(N_0 - dN)$. Hence, for population inversion we require dN > $(N_0 - dN)$, that is, $dN > N_0/2$. Therefore, to obtain population inversion we have to pump more than half the atoms out of the ground state into the upper laser level. This obviously requires a very large amount of energy, which normally only can be delivered by intensive flashlamps. Therefore, three-level lasers are pulsed lasers. One exception of a three-level continuous wave (cw) laser is the argon ion laser. Here, electrical energy is used to ionise the atoms to become the lower laser level.

Much less energy is required to reach the laser threshold in four-level lasers, in which the lower laser level is empty before the pumping process starts.

Although the threshold for population inversion is very high in a three-level system, they can be quite efficient once this threshold is overcome. Ruby lasers pumped by bright flashlamps actually give very high output pulse energies of 20 J easily, and are used in dermatologic applications.

Continuous lasers, like the Nd:YAG laser, normally use four-level systems. Atoms are pumped from the ground state to level 3, which can be composed of different excited states. From there, the upper laser level 2 is populated by fast transition, τ_{32} . This level is long-lived. It is a metastable system because the transition directly to the ground state, τ_{20} , is forbidden. The laser transition occurs to the lower laser level 1 which is short-lived, with fast transition to the ground state. This allows for very efficient pumping.

1.5 Optical Resonators

The optical resonator is composed of two mirrors with the laser active medium located between them. The mirrors might either be flat (Fabry-Perot-Resonator) or curved. In case of plane-parallel mirrors, the reflected beam in the resonator is a plane wave. Because mirrors are never ideal plane, this resonator has losses and is slightly unstable. Stable resonator configurations use concave mirrors and are easier to adjust. Oft-used resonator configurations are the symmetric confocal resonator ($R_1 = R_2 = L$), which has the most stable design and least diffraction losses, as well as the spherical type ($R_1 = R_2 = L/2$).

The relation that describes the stability of an optical resonator is determined by the resonator length, L, and the curvatures of the mirrors, R_1 and R_2 . With the mirror parameters



Fig. 1.11 Different stable and unstable resonator configurations

 $g_1 = 1 - L/R_1$ and $g_2 = 1 - L/R_2$ the stability condition is (Fig. 1.11):

 $0\pounds g_1g_2 \ge 1.$

The resonator geometry determines the form of the output beam with regard to diameter, divergence, and composition of modes. When we consider a stable Gaussian resonator, the multiple reflected beam is always running back in to itself and the phase of the wave front must match the curvature of the mirrors. Because of the marginal condition of the electromagnetic field on the mirrors, a standing wave results along the axis. Therefore, the standing wave spectrum must fulfil the condition:

 $n.\lambda/2 = L$ withintegern = 1, 2, 3, 4, ..., wavelength λ and resonator length L.

This results in an axial spectrum of longitudinal modes with frequency spacing:

 $\Delta n = c/2L.$

As an example, for a resonator length of 0.5 m, neighboring longitudinal modes are separated by $\Delta v = 300$ MHz. A laser emits a characteristic mean wavelength according to the atomic resonance frequency, which is broadened due to the Doppler effect. If pumping of the laser exceeds the threshold because of

losses in the resonator, then laser radiation is emitted and the number of longitudinal modes (axial eigenfrequencies) is determined by the ratio of the Doppler width, Δv_D , to the separation of the modes, c/2L, as demonstrated in Fig. 1.12.



Fig. 1.12 Origin of the longitudinal (axial) modal spectrum. Laser activity is only observed when the gain within the Doppler width exceeds the losses

Practically, only few resonators are working in the stable Gaussian basic mode. Like all systems characterised by marginal conditions, resonators have an entire spectrum of inherent modes. The Gaussian basic mode (Fig. 1.13) represents the oscillation state with the lowest inherent value. Higher modes have a more complex intensity distribution and the transversal expansion of the beam is larger. To increase the output power of a laser, a large amplification volume has to be generated with the assistance of higher transversal modes.

Resonators using spherical mirrors have an output beam with allowed transversal intensity distributions mathematically described by the Hermite-Gaussian modes, the so-called TEM_{nm} modes (transversal electromagnetic modes). Lasers are often designed for the operation in a single transversal mode, the basic Gaussian mode TEM_{00} , which has the smallest beam diameter and the smallest spot size in the focus. Normally they have cylindrical symmetry. In this case the transverse mode patterns are described by a combination of a Gaussian beam profile with a Laguerre polynomial. The modes are denoted TEM_{pl} , where *p* and *l* are integers representing the radial and angular mode orders, respectively.

With p = 1 = 0, the TEM₀₀ mode is the lowest order, or the fundamental transverse mode of the laser resonator, and has the form of a Gaussian beam (Fig. 1.14). The pattern is a round spot with a constant phase. Modes with increasing *p* show concentric rings of intensity, and modes with increasing 1 show angularly distributed lobes. In general there are 2l(p + 1) spots in the mode pattern. The overall size of the mode is determined by the Gaussian beam radius *w*, and this may increase or decrease with the propagation of the beam. During propagation the modes preserve their general shape. Because higher-order modes are larger compared to the TEM₀₀ mode, the fundamental Gaussian mode of a laser may be selected by placing an appropriately sized aperture in the laser cavity that only the fundamental mode can pass through.



Fig. 1.13 Cylindrical transverse mode patterns TEM $_{pl}$. TEM₀₁* represents a special case, the so-called *doughnut mode*, an overlay of the patterns with l = 1, 2, 3

The symmetry of optical resonator often is restricted in lasers by polarising elements like Brewster's angle windows, leading to transverse modes with rectangular symmetry. These modes are named TEM $_{mn}$, with m and n being the horizontal and vertical orders of the pattern. The intensities of the patterns are determined by the order of the Hermite polynomials and demonstrated in Fig.

1.14. The TEM₀₀ mode corresponds to the same fundamental mode as in the cylindrical geometry. The number of spots in the intensity patterns is described by the expression (m + 1)(n + 1). The polarisation of the phase in each spot is offset by π radians with respect to the horizontal or vertical neighbours.



Fig. 1.14 Rectangular transverse mode pattern TEM mn of lasers with Brewster's angle elements

1.6 Laser Radiation

Laser radiation is characterised by its parallelism, the spatial and temporal coherence; its monochromatism, the spatial distribution and divergence (beam quality); and the polarisation, power, and operating mode (pulsed or cw).

The intensity distribution across a Gaussian beam is a bell-shaped Gaussian function. It is a beam of best quality that can be focused to the smallest spot. Lasers emitting in TEM_{00} mode have such a quality. The Gaussian function of the intensity I(r,z) depends on the distance, z, from the beam waist and the radius, w, according to the formula:

$$I(\tau, z) = I_0 (W_0 / W(z))^2 \exp^{-2\tau^2 / W^2(Z)}$$

The intensity at the beam waist, I_0 , can be described in relation to the laser power, P, by the expression:

$$I_0 = 2P/\pi w_0^2 (W/cm^2).$$

With distance, z, from the beam waist at z_0 , the cross- section and the radius, w, of the beam increases and can be expressed as

$$w(z) = w_0 (1 + (z/z_0)^2)^{1/2}.$$

The distance, z, where the cross-section of the beam is twice the area of the focal spot is called the *Rayleigh length* (see Fig. 1.15).

Rayleighlength : $z_0 = p w_0^2 / \lambda$,

where the beam has widened to:



Fig. 1.15 Focused Gaussian beam with spherical wave fronts and the angle of divergence Θ . The radius of the beam waist is w_0 and at a certain distance from the beam waist, w(z). The distance until $\pm z_0$ is twice the Rayleigh length, b

 $w(\pm z_0) = w_0 \sqrt{2}.$

The length between the spots $\pm z_0$, denominated as *b*, is twice the Rayleigh length and is called the *confocal parameter*. The intensity there is half the intensity at the focal spot.

The quality of multimode laser beams, which have no Gaussian profile, is expressed as M^2 , which is independent of λ . It is the relation of the beam parameter product (BPP), which is the product of a laser beam's divergence angle (half-angle) and the radius of the beam at its narrowest point (the beam waist), to that of a Gaussian beam. The BPP quantifies the quality of a laser beam and how well it can be focused to a small spot. M^2 of a Gaussian beam is exactly one. The smaller M^2 of a laser beam is, the smaller is its focal spot.

 $M^2 = w\Theta/(l/p)$ with the divergence of a Gaussian beam,

 $\Theta_0 = \lambda / \pi w_0.$

If the waist of a laser beam is equal to that of a Gaussian beam, then $M^2 = \Theta/\Theta_0$.

Important laser parameters are summarised in Table 1.2.

Wavelength λ (nm or μ m)	Photon energy
Power $P(W)$	For cw lasers
Energy $E = Pt$ (J)	For cw lasers
Peak power P _{max} (W)	For pulsed lasers
Energy E per pulse (J)	For pulsed lasers
Pulse duration <i>t</i> [fs (10^{-15}) to ms (10^{-3})]	For pulsed lasers
Effect on tissue	
Energy density E/A (radiant exposure) (J/cm ²) ($A =$ effective area)	For pulsed lasers
Power density W/A (irradiance) (W/cm ²) ($A =$ effective area)	For cw lasers
Beam quality $L_{\rm w}$ =1/ M^2 (mm mrad ⁻¹)	General for lasers

Table 1.2 Important laser parameters

Polarisation of the laser beam can be achieved by introducing polarising optical components into the resonator, such as deflecting mirrors in folded resonators or by active laser crystals where the faces have Brewster angles (Fig. 1.16). Such a Brewster resonator produces linear polarised light. The disadvantage may be that the losses of the resonator increase and the output power is reduced.



Fig. 1.16 Brewster resonator producing linear polarised laser light [8]

The *operational modes* of lasers are cw, pulsed as interrupted radiation (in ms), pulsed free running (in hundreds of μ s), Q-switched (in ns), or mode-locked (in fs).

Interrupted radiation of a cw laser is done by mechanical or electronic switching with modification of the pulse length. The pulse frequency is low to

moderate, up to 100 Hz. Flashlamp-pumped solid-state lasers in the free-running mode have pulse lengths of 50 μ s up to several hundred microseconds. Pulses of medical dye lasers systems can vary from microseconds to 50 ms.

Shorter pulses with very high intensities in the nanosecond range are produced by Q-switching of the laser. A Q-switch can be introduced into the resonator that reduces the performance factor of the laser resonator. Therefore, the threshold for lasing is very high for a short period of time. During this time, with continuous pumping, pumping power, P_P , the inversion of the number of excited atoms, can be built up, $\Delta N(t)$ (Fig. 1.17). When the Q-switch opens at t =0, the threshold power, S(t), goes down immediately and the amplification is now much higher than the resonator losses. A strong laser pulse is generated with complete depletion of the energy reservoir [21].



Fig. 1.17 Scheme of a Q-switch: P_P is pumping power, S(t) is threshold power, $\Delta N(t)$ is the inversion, and P_L the laser power

The power of the pulse can be estimated according to the expression: $P(t) = h\nu T(c/2L)Vn(t)(W),$

where *h* is Planck's elementary quantum of action (Planck's constant), *v* is the frequency of the laser transition, T = 1 - R is the transmission of the output mirror, *V* is the resonator volume, n(t) is the photon number density, and *L* is the resonator length.

Some typical representatives of Q-switches are shown in Table 1.3.

Table 1.3 Typical representatives of Q-switch devices

Type of Q-switch	Switching time
Rotating aperture in resonator	>10 µs
Rotating mirror or prism	<0.1 µs
FTIR Q-switch	<50 ns
Electro-optical switch	<10 ns

Type of Q-switch	Switching time
Acousto-optical switch	<50 ns
Saturating absorber	<0.1 ns

FTIR frustrated total internal reflection

Ultrashort laser pulses are generated by mode-coupling due to the coherence properties of the laser. Compared to Q-switching, where the shortest pulse durations are in the range of the resonator period, mode-coupling can generate even shorter laser pulses. Mode-coupling finally is attained by forcing a fixed phase relationship, $\Phi_{q+1} - \Phi_q = \Phi$, between the longitudinal laser modes having the distance $\Delta v_{q,q+1} = c/2L$ (see Fig. 1.12). This is associated with interference effects with amplitudes that are larger than those of individual modes. For more information see [4].

1.7 Medical Lasers

Since the invention and commercial availability of lasers, such systems have been widely used for medical applications in the treatment of soft and hard biological tissue, from the Ultraviolet (UV) Excimer laser, ArF at 193 nm, in ophthalmology up to the CO_2 laser in the IR-region at 10,640 nm for surgical interventions. The whole spectrum of lasers is shown in Fig. 1.18.



Fig. 1.18 Medical lasers

1.7.1 Solid-State Lasers 1.7.1.1 Nd:YAG Laser The most important solid-state laser is the Nd: YAG laser, with neodymium as laser ion (Nd^{3+}) doped with 1–1.5% into an yttrium-aluminium-garnet crystal $(Y_3Al_5O_{12})$. Its fundamental wavelength is 1064.2 nm. Frequency doubling with a KTP crystal (potassium titanium oxide phosphate, KTiOPO₄) leads to 532 nm (green light). The third and fourth harmonics are 355 and 266 nm, respectively. The Nd: YAG laser is a four-level system (Fig. 1.10). Other, weaker emission wavelengths are 1,320 and 1058.2 nm (Nd:CaWO₄). The overall efficiency of a flashlamp-pumped Nd: YAG laser is less than 2%. Therefore, it must be cooled with water. Diode pumped systems (805-nm laser diodes) reach efficiencies of about 15% at output power of about 10–15 W. Nd:YAG lasers are used as cw lasers up to 100 W for surgical applications, e.g., long pulse lasers in dermatology or Q-switch lasers with-to 20-ns pulse length in urology (lithotripsy) and dermatology (tattoo removal).

1.7.1.2 KTP Laser

KTP lasers are frequency-doubled Nd:YAG lasers using KTP crystals inside the cavity or externally with an emission wavelength of 532 nm in the green part of the visible spectrum.

1.7.1.3 Er:YAG Laser

The Er:YAG laser with a 2.94-µm wavelength is well suited for ablation of soft and hard biological tissue because the wavelength is highly absorbed in water (μ $_a = 10^4$ cm⁻¹). Erbium ions are doped in high concentration (~50%) in YAG or other crystals (Er:YSGG: 2.78 µm; Cr:Er:YSGG: 2.80 µm; Cr:Tm:Er:YAG: 2.64 µm). This laser exists only as pulsed laser system because the rather complex energy-level scheme of erbium with important energy-transfer upconversion (ETU; see Fig. 1.19) and cross-relaxation (CR) processes. Laser transition occurs from the energy level ${}^{4}I_{11/2}$ to ${}^{4}I_{13/2}$. The lower laser level has a longer lifetime of 9 ms compared to the upper laser level. Only by hard pumping and the efficient upconversion processes is lasing possible. The ETU process (${}^{4}I_{13/2}$, ${}^{4}I_{13/2}$) ® (${}^{4}I_{15/2}$, ${}^{4}I_{9/2}$) leads to a fast depletion of the lower laser level. The other advantage of the ETU process is that half of the ions undergoing this process are upconverted to the ${}^{4}I_{9/2}$ upper level and, by subsequent multiphonon relaxation, are recycled to the ${}^{4}I_{11/2}$ upper laser level, from where they can each emit a second laser photon [18, 19].



Fig. 1.19 Energy level scheme of the Er:YAG laser. *GSA* ground state absorption; *ETU* energy-transfer upconversion between neighboring erbium ions; *CR* cross-relaxation

Pulsed Er: YAG laser systems (pulse lengths from 100 to 500 μ s) have a mean power of up to 30 W and repletion rates of up to 50 Hz, with efficiencies of 1– 2%. They are used in dermatological applications for superficial tissue ablation, in dentistry, and in minimal invasive surgery. The laser light is transmitted either by an articulated arm or by fibres. Germanium oxide fibres proved to be stable in clinical routine work.

1.7.1.4 Holmium: YAG Laser

Flashlamp- or diode-pumped Ho:YAG lasers can be operated in the pulsed or the cw mode. In the pulsed mode, with pulse durations $200-600 \ \mu s$ and energies from 0.2 to 3.5 J, repetition rates up to 50 Hz have a mean power of $30-40 \ W$. Combining several laser heads, high-power systems are also available. The wavelength is 2,080 nm.

Applications with low power first appeared in ophthalmology for refractive surgery procedures called laser thermal keratoplasty to correct mild to moderate cases of hyperopia and astigmatism. In recent years applications in urology and endourology are more important. Besides stone disintegration, treatment of benign prostate hyperplasia (BPH) by enucleation or vaporisation has become a standard surgical intervention [11, 13].

The most efficient laser material for high-power flashlamp-pumped lasers is Cr,Tm,Ho:YAG. The laser's active ion is Ho^{3+} with a concentration of about 0.5% atomic units (AU) in the host material YAG, co-doped with thulium ions (Tm³⁺, 6% AU) and for lamp pumping with chromium ions (Cr³⁺, 1% AU). The scheme of the lamp-pumped holmium system is shown in Fig. 1.20.



Fig. 1.20 Scheme of the Cr,Tm,Ho:YAG laser system

The pump light in Fig. 1.20 is absorbed by the broad absorption bands $A_2 T_2$ of the chromium ions at wavelengths of 420 and 600 nm or directly by energy levels of the Tm³⁺ ions from ³H₆ ³F₄ by diode pumping. From all the excited levels, rapid radiationless energy transfer occurs to the energy level ³F₄ of the holmium ions. The highly excited Tm³⁺ ions are deactivated from the ³F₄ to the ³H₄ level. The energy difference of the levels ³F₄ and ³H₄ fit very well with the energy difference between the low energy level ³H₆ and the ³H₄ level of the Tm³⁺ ion. So the energy of the deactivated Tm³⁺ ion is, with a high probability, resonantly transferred to a neighboring Tm³⁺ ion, which then is excited from the low energy level ³H₆ to the ³H₄ level. This resonant energy transfer is called

cross-relaxation with quantum efficiency raised to *two*. One pump photon can excite two Tm³⁺ ions to the ³H₄ level by this cross-relaxation. The ³H₄ excitation energy migrates between neighbouring Tm³⁺ ions via energy resonant processes. If a holmium ion is reached, the energy can be transferred without radiation to the quasi-resonant upper laser level ⁵I₇ of the Ho³⁺ ion. The laser radiation terminates at the ⁵I₈ level of the Ho³⁺ ion.

The lower laser level of the Ho³⁺ ion has only a small energy difference to the ground state and is therefore highly populated at ambient temperatures. Holmium laser systems operating at room temperature are quasi-three-level laser systems. Consequently, a very high excitation energy is necessary to reach a population inversion. The resulting thermal load to the laser material in lamp-pumped systems allows laser operation only in the pulsed mode. The laser output is very temperature-dependent. It will be reduced by about 3% per degree of rising temperature.

1.7.1.5 Thulium Laser (Tm:YAG)

Laser therapy of symptomatic BPH is a promising, less invasive alternative to the traditional BPH surgery. Besides the holmium laser, the cw thulium laser at a 2- μ m wavelength is used as well [6]. Compared to the KTP laser, the thulium laser is more efficient [25]. The absorption coefficients of the mid-IR lasers are compared in Fig. 1.21. Tm: YAG lasers have a slightly higher absorption in water than the Ho: YAG lasers. Therefore, Tm: YAG lasers are well suited for vaporisation of the prostatic tissue.



Fig. 1.21 Absorption coefficients and penetration depths in water of mid-infrared laser [7]

1.7.1.6 Alexandrite Laser

This solid-state laser is tunable in the wavelength range 700–830 nm, normally being operated at 750 nm. Alexandrite laser medium is a chromium-doped chrysoberyl (Cr:BeAl₂O₄, gemstone alexandrite) and can operate in cw (up to 100 W) or pulsed mode. Because light of this wavelength region is absorbed by melanin and dyes, but not significantly by blood, it can be used for the destruction of melanin-containing structures (hair roots, pigmented lesions). The first applications were reported in the fragmentation of kidney stones [17].

Also not a true four-level laser, the vibronic transition on which this laser usually operates permits operation similar to a four-level laser. The scheme of the energy levels is shown in Fig. 1.22.



Fig. 1.22 Scheme of the energy levels of the alexandrite laser

The four-level laser depicted in Fig. 1.22 has a ground manifold, a pump manifold, and an upper laser manifold, similar to a three-level laser. But, in contrast to the three-level system, there is a fourth manifold, the lower laser manifold, in which the lower laser level resides, well above the ground manifold. Because the population density of the ground level does not have to be overcome, $Cr:BeAl_2O_4$ operation on a vibronic transition resembles the operation of a four-level laser with a low laser threshold.

1.7.1.7 Ruby Laser ($Cr^{3+}:Al_2O_3$)

Although the ruby laser was the first laser, which was developed by Maiman in 1961, it was for a long time ignored and not used in medical applications. The emission wavelength is 694 nm. The laser is a three-level laser and therefore more than 50% of the ions in the ground state must be excited to get an inversion. Only strong pumping by powerful flashlamps can produce this inversion. Modern laser systems with advanced technology are able to produce high pulse energies up to 20 J. The laser light is transmitted through fibres or articulated arms. The deep red color of the laser light is absorbed by melanin and dark pigments. Therefore, the medical applications are mostly in dermatology, similar to the alexandrite laser, to remove hairs or tattoos with black or blue color.

The laser active ion is Cr^{3+} replacing Al ions, Al³⁺, in the host material, sapphire (Al₂O₃), at a doping concentration of 0.05% AU of Cr_2O_3 . The two strong absorption bands in the ruby crystal are around 400 and 550 nm. Optical pumping occurs from the ground state, ⁴A₂, to the two excited states of the chromium ion, ⁴T₁ and ⁴T₂, from where, quickly and without radiation, the energy is transferred to the upper laser level, *E*. Laser emission of 694 nm occurs from energy level *E* to the ground state ⁴A₂.

1.7.1.8 Titanium-Sapphire Laser (Ti:Al₂O₃)

The Ti:sapphire laser was developed by Moulton and first reported by him in 1986 [15]. This solid-state laser is pumped by a frequency-doubled Nd:YAG laser and can be tuned between 660 and 1,160 nm. With mode-locking the laser emits ultra-short femtosecond pulses (45–180 fs) at high repetition rates (80–100 MHz) or picosecond pulses. Due to the high power of the laser pulses, Ti:sapphire lasers are used as a light source in two-photon microscopy or in ophthalmology for cutting transparent ocular tissue (flap preparation of the cornea for vision corrections).

1.7.2 Gas Lasers

1.7.2.1 Carbon Dioxide Laser

The carbon dioxide (CO2) laser is still one of the important lasers used for medical surgery and industrial applications. With a power range from mW to tens of kW in the cw mode and an efficiency of up to 30%, laser systems are compact and economic. The sealed laser tube is filled with a mixture of gases, CO_2 (1–9%), N₂ (13–45%), and helium (60–85%). Excitation occurs by direct current high-voltage gas discharge or through high frequency (RF). It must be mentioned

that only with direct current supply units is the superpulse technique possible; it generates pulses <1 ms and with a peak power about ten times higher than the mean power. The gas components have different reaction mechanisms during the laser process.

The structure of the CO_2 molecule is linear, with the carbon atom in the centre. Such molecule configurations can vibrate in symmetric and asymmetric stretch modes as well as in bending modes (Fig. 1.23). The energies associated with molecular vibration are quantised just like electron energies; therefore, only certain vibrational levels are possible. The possible forms of resonant vibration are referred to as the vibrational modes of a molecule.



Fig. 1.23 Scheme of the energy levels of a CO₂ laser and the vibration modes of the linear CO₂ molecule

During gas discharge free electrons collide with N_2 molecules and excite them. The CO₂ is excited through a collision with excited N_2 molecules. This transfer of energy occurs by a resonant effect. Because the vibrational energy levels of N_2 are metastable and have an energy very close to that of the first energy level of the asymmetric stretch mode of CO₂, they have ample time to transfer their energy and excite the CO_2 molecules. The lasing occurs when CO_2 in the excited asymmetric mode makes the transition to the bending or symmetric stretch modes. The CO_2 then returns to its ground state through another collisional transfer of energy with the helium atoms. With this lasing process the CO_2 laser is a three-level laser with direct pumping into the upper laser level because the lower laser level is depleted very quickly through the presence of helium in high concentration. One receives a bunch of laser lines around the central laser wavelengths at 10.6 and 9.6 μ m.

A special variant of the CO_2 laser is the TEA (*T*ransversely *E*xcited *A*tmospheric Pressure) CO_2 laser. High peak powers can only be achieved by increasing the density of the excited CO_2 molecules. To avoid too high voltage necessary to achieve gas breakdown in longitudinal configuration, Beaulieu invented the TEA-laser in 1970 [2]. Mechanical Q-switching of CO_2 lasers to generate <1 µs pulses at kHz repetition rates can be used to cut bone tissue [12].

Transmission of the CO₂ laser light usually is provided by an articulated arm with mirrors. A new fibre technology invented at Massachusetts Institute of Technology, Cambridge, MA, USA, and developed by OmniGuide Inc., a wavelength-scalable hollow optical fibre from 750 nm to 10.6 μ m with large photonic bandgaps [23] opens the possibility of flexible CO₂ laser transmission and endoscopic applications.

1.7.2.2 Argon Ion Laser

Noble gas ion lasers are cw lasers. Prominent representatives are the Ar^+ laser and Kr^+ laser. The wavelengths of the Argon ion laser are 363,488 and 514.5 nm in the UV, blue, and green spectral regions. Intense lines of the Krypton ion laser are at 530.9nm, 568.2 nm, and in the red region are at 647 and 676.4 nm. The output power in multi-line mode of the Ar^+ laser is between 2 and 100 W; the Kr^+ laser has about 20 W. Only in the UV spectrum is the Ar^+ laser still unrivaled.

Ion lasers are very inefficient: about one-tenth of a percent of electrical current is converted to laser light. They need high electrical current supply and water cooling. Nevertheless, the Ar⁺ laser has been used to pump Ti-sapphire and dye lasers. It mostly has been replaced by more efficient frequency-doubled Nd:YAG lasers.

The argon laser consits of a plasma tube filled with argon gas at low pressure, 0.01-0.1 mbar. Plasma discharge current up to 60 A ionises the argon from its Ar ground state, $3p^6$, by successive electron collisions to the ionised Ar⁺ ground

level, $3p^5$. Further electron collisions excite the argon ion to the Ar⁺ $3p^44p$ (upper laser) levels. Laser transition occurs from these levels to the lower laser levels, Ar^+3p^44s .

1.7.2.3 Helium-Neon (HeNe) Laser

The best known cw gas laser is the HeNe laser. The active medium is a mixture of helium and neon in proportions of 5:1-10:1 at low pressure (1 mbar) in a glass capillary. The actual laser transition occurs in the neon. The gas discharge tube operates at 1.5-2 kV, maintaining a glow discharge or plasma. Electron collisions excite helium atoms from the ground level to the 2^1S_0 and 2^3S_1 state. The energy of the helium atoms is transferred by collision with the neon atom to the neighboring energy levels of the neon, $3s_2$ and 2s levels. The $3s_2$ level of neon is an example of a metastable atomic state, meaning that it is only after a relatively long period of time – on atomic time scales – that the Ne*($3s_2$) atom de-excites to the $2p_4$ level by emitting a photon with a 632.8-nm wavelength. The excited Ne* ($2p_4$) atom rapidly de-excites to its ground state by emitting additional photons or by collisions with the plasma tube walls. Because of the extreme quickness of the de-excitation process, at any time in the HeNe plasma there are more neon atoms in the $3s_2$ state than there are in the $2p_4$ state, and a population inversion is said to be established between these two energy levels.

Further laser lines of the HeNe laser are at 543, 594, 604, and 612 nm in the visible spectrum and at 1,152 and 3,391 nm in the IR spectrum. HeNe lasers are often used as aiming beams, for spectroscopic or diagnostic applications and biostimulation. Maximum output power is 100 mW; small devices have 1–5 mW output power.

1.7.2.4 Excimer Laser

Basov and his colleagues in Moscow first discovered in 1971 the stimulated UV emission of an excited xenon dimer (Xe₂) at 176 nm [1]. Excimer is a short form of the expression "excited dimers." Molecules, such as noble gas halides (ArF, KrF, XeCl, XeF), are stable only in their excited states and not in their ground states. The laser medium consists of such molecules in a buffer gas like helium or neon at a total pressure of 2-3 bar.

Figure 1.24 shows the potential curves of for a halogen rare gas molecule, (RGH) in its excited state and ground state. The electronically excited RGH (XY)* forms the upper laser level with a lifetime of several nanoseconds. The laser radiation occurs during transition from this excited state to the non excited

state of rare gas and halogenide. The efficiency of excimer lasers is about 2%. They exist only as pulsed lasers, with pulse widths ranging from several ns to 100 ns. Maximum mean power is 200 W. The repetition rate is up to kHz.



Fig. 1.24 Typical potential curves of rare gas-monohalide excimers

The wavelength of an excimer laser depends on the rare gas and the halogenide. The wavelengths mainly used in the UV spectrum are listed in Fig. 1.25.



Fig. 1.25 Laser wavelengths of the different excimers in nanometers

1.7.3 Dye Lasers

The invention of the dye laser in 1966 by Fritz P. Schäfer was a chance discovery. When he directed the beam of a ruby laser through a glass cuvette with fluorescing dye he observed a lasing effect. The reflections of the glass walls were sufficient to react as a resonator. Later, T.W. Hänsch used the dye laser with frequency-selective optical elements for spectroscopy. Today, dye lasers are still in use as cw lasers pumped with argon ion lasers or frequency-doubled Nd:YAG lasers. The advantage of the dye laser is its broad tuning range, up to 100 nm with one dye. Numerous dyes are available, covering the whole range in the visible spectrum. However, low-cost diode lasers have become an interesting alternative, specially for photodynamic therapy. Pulsed dye lasers are routinely used in dermatology for treatment of vascular malformations.

Some of the fluorescing dyes used in dye lasers are rhodamine 6G (R6G), fluorescein, coumarin, stibene, umbelliferone, tetracene, and malachite green [16].

The dye laser is a four-level laser. R6G has been chosen as an example in Fig. 1.26. Photon absorption of the R6G molecules excites the molecules from the bottom of the S₀-band to the higher S-bands. All excited states in the higher S-bands contribute to the laser activity because there is a very fast transition (10^{-12} s) of the vibration-rotation energy levels within one S-band to the lowest energy level, where they will have a Boltzmann distribution. Also, transitions from higher S-bands, to the S₁-band, as demonstrated in Fig. 1.26, are very fast (10^{-11} s) compared to the fluorescence lifetime of $\geq 10^{-8} \text{ s}$.



Fig. 1.26 Scheme of a dye laser. Inserts show the structure of rhodamine 6G (*bottom*) with absorption and emission spectra (*top*)

Transitions from S_1 to the vibration-rotational energy levels in the S_0 -band are responsible for the different laser lines that continuously can be tuned within the range of the dye used.

1.7.4 Diode Laser (Semiconductor Laser)

Hall and his team at General Electric were the first to demonstrate the emission of a semiconductor laser diode in 1962 [9].

Laser diodes are formed by doping very thin layers on the surface of a crystal wafer to produce an *n*-type region (negative, increase of negative free charges – electrons) and *p*-type region (positive, increase of free positive charges – holes), one above the other, resulting in a p-n junction. Applying a forward electrical bias causes the two species of charge carriers, holes, and electrons, to be injected from both sides of the p-n junction into the depletion zone, situated in the centre of the p-n junction. Holes are injected from the p-doped semiconductor and electrons from the n-doped semiconductor (see Fig. 1.27). Recombination or

annihilation of both charges results in a spontaneous emission of a photon, with energy equal to the difference between the electron and hole states involved.



Voltage applied



The difference between the photon-emitting semiconductor laser and conventional diodes lies in the use of a different type of semiconductor. Photonemitting semiconductors are the so-called *direct bandgap* semiconductors. These semiconductors are compound semiconductors, in contrast to silicon and germanium, which are single-element semiconductors that do not emit photons. Compound semiconductors have crystalline Structures that are virtually-identical to silicon or germanium but use alternating arrangements of two different atomic species in a checkerboard-like pattern. Such materials are gallium arsenide, indium phosphide, gallium antimonide and gallium nitride.

In the absence of lasing, holes and electrons can coexist in proximity to one another, without recombining, for a certain time, termed the "upper-state lifetime" or "recombination time," about a nanosecond, before they recombine. A nearby photon with energy equal to the recombination energy can cause recombination by stimulated emission. This generates another photon of the same frequency, travelling in the same direction, with the same polarisation and phase as the first photon. This stimulating emission causes light amplification in the p-n junction region proportional to the number of electrons and holes being injected.

A simple form of a laser diode consists of an optical waveguide as the optical cavity, made on the crystal surface, such that the light is confined to a relatively narrow line (see Fig. 1.28). When the two ends of the crystal are cleaved to form perfectly smooth, parallel edges, they form a Fabry-Perot resonator. The laser diode will start to lase when the losses (spotaneous emission, absorption, incomplete reflection) are surpassed by the stimulating emission.



Fig. 1.28 Schematic drawing of a laser diode with asymmetric light emission. Emitting surface is about 1 μ m high and 100 μ m wide

Laser diodes are very efficient, with an electrical current to light conversion of up to 60%. Therefore, they are well suited for medical applications because they are very powerful (up to 100 W), they are small devices, and they are flexible in wavelength. Laser diodes are available from the UV spectrum (for chromophore excitation) into the visible spectrum until the near-IR spectrum for surgical applications.

Laser diodes with a single emitter, as demonstrated in Fig. 1.28, may have an optical output of 10 to 40 W. Peltier elements are normally integrated to cool the diodes. Even higher power is obtained by diode bars, where 20–50 single emitters, 100 μ m broad, are aligned in one stripe with a fill factor of 50%. Such bars can be mounted one above the other to form a stack of several 100 W output power.

1.8 Take Home Pearls

- New developments in laser technology provide higher efficiency in laser light generation and miniaturisation to improve the handling in medical applications.
- Laser light is now available in the range of UV to far IR.
- According to the needs of specific applications, the radiation of laser light can be modulated from cw to femtosecond pulses.
- Experienced dermatologists have now the possibility to select the optimal laser wavelength and pulse regime.

References

- Basov NG, Danilychev VA, Popov YM. Stimulated emission in the vacuum ultraviolet region. Sov J Quantum Electron. 1971;1:18–22.
 [CrossRef]
- Beaulieu AJ. Transversely excited atmospheric pressure CO₂ lasers. Appl Phys Lett. 1970;16:504–5.
 [CrossRef]
- 3. Demtröder W. Atoms, molecules and photons. New York: Springer; 2006.
- 4. Diels JC, Rudolph W. Ultrashort laser pulse phenomena. New York: Academic Press; 2006.
- 5. Einstein A. Zur Quantentheorie der Strahlung. Phys Zeitschr. 1917;18:121-8.
- 6. Elhilali M, Elzayat EA. Laser prostatic surgery: an update. Afr J Urol. 2008;14:1–14. [CrossRef]
- 7. Frank F, Wondrazek F. Erbium:YAG laser. In: Berlien HP, Müller GJ, editors. Applied laser medicine. Berlin: Springer; 2003.
- Gross H. Handbook of optical systems, vol. 4. (Materials) New York: Wiley; 2005. [CrossRef]
- Hall RN, Fenner GE, Kingsley JD, Soltys TJ, Carlson RO. Coherent light emission from GaAs junctions. Phys Rev Lett. 1962;9:366–9.
 [CrossRef]
- 10. Hecht J. Laser pioneers. New York: Academic Press; 1992.
- Hoffmann RM. Laser prostatectomy versus transurethral resection for treating benign prostatic obstruction: a systematic review. J Urol. 2003;169:210–5.
 [CrossRef]
- Ivanenko MM, Hering P. Wet bone ablation with mechanically Q-switched high-repetition-rate CO₂ laser. Appl Phys B. 1998;67:395–7. [CrossRef]